

Office of Weights and Measures
Metrology Laboratory

Office: 118 West Capitol Avenue, Pierre, SD 57501
 Lab: 1100 Otter Rd, Bldg D, Sturgis, SD 57785
 Lab: 605-347-7541, Office: 605-773-3697, Cell: 605-280-4572
 Email: ron.peterson@state.sd.us <https://dps.sd.gov/inspections/weights-measures>

CALIBRATION CERTIFICATE

Capital Scale	SA# 61	Certificate number: M25009
Physical Address:	Billing Address:	
3021 Valley Forge St	3021 Valley Forge St	
Bismarck, ND 58503	Bismarck, ND 58503	
Contact: Travis Will		Received Date: 09/30/2024
Phone: 701-255-1556		Certificate Issued: 10/01/2024

Artifacts Submitted and Summary of Results:

Quantity	Artifact	Total Pieces	Recvd in Tol	Adjusted	Rejected	As Left In Tolerance
2	2000 lb weight carts	2	1	1	0	2
16	1000 lb weights	16	16	4	0	16
20	50 lb weights	20	0	20	0	20
1	Metric kit	14	14	0	0	14
1	Avoirdupois kit	20	20	0	0	20

Uncertainty Statement: The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 2008 ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. For weight carts, factors included on the inspection checklist have not been included in the calibration uncertainty. However, factors on the checklist may contribute measurement errors that are significant if not properly maintained during use.

Conformity Statement: The artifacts submitted for this calibration are calibrated to NIST Handbook 105-1 (1990 or 2019), NIST Handbook 105-8 (2019), NIST Handbook 105-3 (2010), or ASTM E617 (2023), Standard Specification for Laboratory Weights and Precision Mass Standards specifications. The reported test values relate only to the observations made at the time and conditions of the test. Artifacts fully comply with all requirements (both specifications and tolerances) of the applicable documentary standard unless otherwise stated. Stated expanded uncertainties are less than one-third of the specified tolerances (maximum permissible errors, m.p.e.) for mass calibrations and less than the specified tolerances for volume calibrations. The correction value plus or minus the expanded uncertainty is within the stated tolerances. It is the decision rule of the SD State Metrology Laboratory that any cast weights determined to have a correction within 66 % of the upper tolerance or 50 % of the lower tolerance will be adjusted closer to zero mass correction, even if the mass correction originally met the applicable tolerance.

Traceability Statement: The Standards of the SD Metrology Laboratory used for comparison are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory certificate number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

This document does not represent or imply endorsement by NIST Office of Weights and Measures or any agency of the State and/or national governments. This report may not be reproduced, except in full without the written approval of this laboratory. The client must not use this

 Ron E Peterson, Metrologist	10/01/2024	 Dwight R Johnson, Reviewer
10/01/2024		



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: **Capital Scale**
 Calibration Date: **09/30/2024**

Certificate Number: **M25009**

Environmental conditions at time of test:

Temperature: 21.44 °C Humidity: 48.26 % Pressure: 674.52 mmhg

Test method used: SOP 33 Calibrations of Weight Carts, May 2019

Test equipment used: Recently calibrated weights and a Mettler SLS510 Load Cell with IND570 Indicator.
 Vaisala PT301

Condition of Carts: Used but in good condition

Manufacturer: Unk

SN: 16037

Nominal (lb)	AS Found (lb)	As Found (g)	As Left (lb)	As Left (g)	Uncertainty (lb)	k	Tolerance (lb)	Condition as Left
2000	-9.47	-4299.10	-0.07	-30.19	0.12	2.01	0.70	Adjusted

Notes:

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory.

The above weight cart was allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted if needed and as noted above to as close as practical to zero error. All fluid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require recalibration of the weight cart prior to subsequent use.

Conformity Assessment:

The weight cart identified on this calibration certificate complies with NIST Handbook 105-8, 2019 specifications and tolerances. Additional details regarding the assessment are included in the associated checklist that is an integral part of this calibration certificate. The weight cart was found (or adjusted) to within the specified tolerances.

The above weight cart was compared with standards of the State of South Dakota, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and have current calibration values. The assigned certificate number provides documented evidence for measurement traceability.

Ron E Peterson, Metrologist

09/30/2024

Dwight R Johnson, Reviewer

09/30/2024

Ver 20240214



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



Inspection Checklist for Weight Cart

Calibrated for: Capital Scale Certificate number: M25009
Calibration Date: 10/01/2024

Manufacturer: Date of Manufacture:
Model Number: ID/SN Number:

<input checked="" type="checkbox"/>	Nominal Mass of Weight Cart	<input type="text" value="2000 lbs"/>	Suitably marked: Yes/No	<input type="text" value="Yes"/>
<input checked="" type="checkbox"/>	Powered by:	Electric/generator <input checked="" type="checkbox"/>	Diesel <input type="text"/>	Gasoline <input type="text"/>
<input type="checkbox"/>	Fluid Levels:	Engine Oil <input type="text"/>		
		Hydraulic Fluid <input type="text"/>		Sealed: Yes/No <input type="text"/>
		Battery <input checked="" type="checkbox"/>		Sealed: Yes/No <input type="text" value="Yes"/>
		Liquid Fuel <input type="text"/>		Reference Line Present: Yes/No <input type="text"/>
<input checked="" type="checkbox"/>	Fluid drain tubes extend beyond the body of the cart: Yes/No		<input type="text"/>	
<input checked="" type="checkbox"/>	Number of axles:	<input type="text" value="2"/>		
<input checked="" type="checkbox"/>	Number /Size of Tires	<input type="text" value="15x5x11.25"/>		
<input checked="" type="checkbox"/>	Sealed wheel bearings: Yes/No	<input type="text" value="Yes"/>		
<input checked="" type="checkbox"/>	Drain holes present in locations where water may accumulate: Yes/No		<input type="text" value="Yes"/>	
<input checked="" type="checkbox"/>	Weight restraint railing permanently fixed and solid: Yes/No		<input type="text" value="Yes"/>	
<input checked="" type="checkbox"/>	Adjusting cavity accessible: Yes/No	<input type="text" value="Yes"/>	Approximate capacity:(lbs)	<input type="text" value="50"/>
<input checked="" type="checkbox"/>	Adjusting cavity sealed: Yes/No	<input type="text" value="Yes"/>		
<input checked="" type="checkbox"/>	Service brakes functioning properly: Yes/No		<input type="text" value="Yes"/>	
<input checked="" type="checkbox"/>	Parking brakes functioning properly: Yes/No		<input type="text" value="Yes"/>	
<input type="checkbox"/>	Remote control functioning properly: Yes/No		<input type="text"/>	

General condition at time of calibration (note any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals).

List and report any repair and maintenance performed, parts replaced, etc., Leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc. Include any comments or changes since the last calibration.

Ron E Peterson Ron E Peterson, Metrologist 09/30/2024 *Dwight R Johnson* Dwight R Johnson, Reviewer 09/30/2024
Ver 20240214



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: **Capital Scale**
 Calibration Date: **09/30/2024**

Certificate Number: **M25009**

Environmental conditions at time of test:

Temperature: 20.75 °C Humidity: 50.5 % Pressure: 674.6 mmhg

Test method used: SOP 33 Calibrations of Weight Carts, May 2019

Test equipment used: Recently calibrated weights and a Mettler SLS510 Load Cell with IND570 Indicator.
 Vaisala PT301

Condition of Carts: Used but in good condition

Manufacturer: Unk **SN:** 16039

Nominal (lb)	AS Found (lb)	As Found (g)	As Left (lb)	As Left (g)	Uncertainty (lb)	k	Tolerance (lb)	Condition as Left
2000	-0.10	-45.40	-0.10	-45.40	0.12	2.01	0.70	In-Tolerance

Notes:

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory.

The above weight cart was allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted if needed and as noted above to as close as practical to zero error. All fluid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require recalibration of the weight cart prior to subsequent use.

Conformity Assessment:

The weight cart identified on this calibration certificate complies with NIST Handbook 105-8, 2019 specifications and tolerances. Additional details regarding the assessment are included in the associated checklist that is an integral part of this calibration certificate. The weight cart was found (or adjusted) to within the specified tolerances.

The above weight cart was compared with standards of the State of South Dakota, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and have current calibration values. The assigned certificate number provides documented evidence for measurement traceability.

Ron E Peterson, Metrologist
 Ver 20240214

09/30/2024

Dwight R Johnson, Reviewer

09/30/2024



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



Inspection Checklist for Weight Cart

Calibrated for: Capital Scale Certificate number: M25009
 Calibration Date: 10/01/2024

Manufacturer: **Unk** Date of Manufacture: **2016**
 Model Number: **2016-2** ID/SN Number: **16039**

<input checked="" type="checkbox"/>	Nominal Mass of Weight Cart	2000 lbs	Suitably marked: Yes/No	Yes
<input checked="" type="checkbox"/>	Powered by:	Electric/generator <input checked="" type="checkbox"/>	Diesel <input type="checkbox"/>	Gasoline <input type="checkbox"/>
<input type="checkbox"/>	Fluid Levels:	Engine Oil <input type="checkbox"/>		
		Hydraulic Fluid <input type="checkbox"/>		Sealed: Yes/No <input type="checkbox"/>
		Battery <input checked="" type="checkbox"/>		Sealed: Yes/No Yes
		Liquid Fuel <input type="checkbox"/>		Reference Line Present: Yes/No <input type="checkbox"/>
<input checked="" type="checkbox"/>	Fluid drain tubes extend beyond the body of the cart: Yes/No		<input type="checkbox"/>	
<input checked="" type="checkbox"/>	Number of axles:	2		
<input checked="" type="checkbox"/>	Number /Size of Tires	15x5x11.25		
<input checked="" type="checkbox"/>	Sealed wheel bearings: Yes/No	Yes		
<input checked="" type="checkbox"/>	Drain holes present in locations where water may accumulate: Yes/No		Yes	
<input checked="" type="checkbox"/>	Weight restraint railing permanently fixed and solid: Yes/No		Yes	
<input checked="" type="checkbox"/>	Adjusting cavity accessible: Yes/No	Yes	Approximate capacity:(lbs)	50
<input checked="" type="checkbox"/>	Adjusting cavity sealed: Yes/No	Yes		
<input checked="" type="checkbox"/>	Service brakes functioning properly: Yes/No		Yes	
<input checked="" type="checkbox"/>	Parking brakes functioning properly: Yes/No		Yes	
<input type="checkbox"/>	Remote control functioning properly: Yes/No		<input type="checkbox"/>	

General condition at time of calibration (note any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals).

List and report any repair and maintenance performed, parts replaced, etc., Leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc. Include any comments or changes since the last calibration.

Ron E Peterson

Dwight R Johnson

Ron E Peterson, Metrologist 09/30/2024 Dwight R Johnson, Reviewer 09/30/2024
 Ver Ver 20240214



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: **Capital Scale** Certificate number: **M25009**
 Calibration Date: **09/30/2024** Purchase Order Number: **0**

Environmental conditions at time of test: Serial#
Temperature: 21.1 °C **Humidity: 50.8 %** **Pressure: 674.48 mmhg**
Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019
Test equipment used: Lab standards traceable to the SI, an XPE604KMC balance, and a Vaisala PTU301
Condition of Weights: Cleaned and painted

Artifact(s): 16 - 1000 lb weights

Nominal	SN/ID	Correction as Found		Correction as Left		ASTM E 617 Class 6 Tolerance (g)	Uncertainty g	k	Condition As Left
		lb	g	lb	g				
1000 lb	10.2	-0.05	-23.5	0.00	0.0	45	4.7	2.0	Adjusted
1000 lb	10.4	-0.04	-18.6	-0.04	-18.6	45	4.7	2.0	In-Tolerance
1000 lb	10.7	-0.03	-12.6	-0.03	-12.6	45	4.7	2.0	In-Tolerance
1000 lb	13.1	-0.08	-35.3	0.00	0.0	45	4.7	2.0	Adjusted
1000 lb	13.2	-0.06	-27.1	0.00	-0.1	45	4.7	2.0	Adjusted
1000 lb	13.3	-0.03	-15.2	-0.03	-15.2	45	4.7	2.0	In-Tolerance
1000 lb	13.4	-0.02	-7.6	-0.02	-7.6	45	4.7	2.0	In-Tolerance
1000 lb	13.5	-0.03	-11.8	-0.03	-11.8	45	4.7	2.0	In-Tolerance
1000 lb	13.5	-0.02	-7.1	-0.02	-7.1	45	4.7	2.0	In-Tolerance
1000 lb	13.6	-0.04	-19.4	-0.04	-19.4	45	4.7	2.0	In-Tolerance
1000 lb	13.7	-0.08	-37.9	0.00	0.1	45	4.7	2.0	Adjusted
1000 lb	13.9	-0.03	-12.9	-0.03	-12.9	45	4.7	2.0	In-Tolerance
1000 lb	16.1	-0.01	-4.6	-0.01	-4.6	45	4.7	2.0	In-Tolerance
1000 lb	16.3	-0.01	-5.3	-0.01	-5.3	45	4.7	2.0	In-Tolerance
1000 lb	16.5	-0.05	-22.3	-0.05	-22.3	45	4.7	2.0	In-Tolerance
1000 lb	16.6	-0.02	-8.3	-0.02	-8.3	45	4.7	2.0	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist

Dwight R Johnson, Reviewer
09/30/2024
09/30/2024



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale **Certificate number:** M25009
Calibration Date: 10/01/2024 **Purchase Order Number:**

Environmental conditions at time of test: Serial#

Temperature: 21.5 °C **Humidity:** 48.2 % **Pressure:** 674 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR50035C, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Cleaned and painted

Artifact(s): **20 50 lb weights**

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	k	Condition As Left
50 lb	0	-5243	-3	2300	200	2.02	Adjusted
50 lb	1	-9663	-3	2300	200	2.02	Adjusted
50 lb	3	-9373	7	2300	200	2.02	Adjusted
50 lb	7	-6853	-3	2300	200	2.02	Adjusted
50 lb	11	-9053	2	2300	200	2.02	Adjusted
50 lb	12	-6148	-3	2300	200	2.02	Adjusted
50 lb	17	-129703	2	2300	200	2.02	Adjusted
50 lb	28	-6173	2	2300	200	2.02	Adjusted
50 lb	38	-6488	-3	2300	200	2.02	Adjusted
50 lb	56	-2353	2	2300	200	2.02	Adjusted
50 lb	59	-9398	-8	2300	200	2.02	Adjusted
50 lb	65	-8588	-8	2300	200	2.02	Adjusted
50 lb	67	-9213	-3	2300	200	2.02	Adjusted
50 lb	68	-10108	2	2300	200	2.02	Adjusted
50 lb	78	-3263	-3	2300	200	2.02	Adjusted
50 lb	79	-9468	2	2300	200	2.02	Adjusted
50 lb	U	-10003	-3	2300	200	2.02	Adjusted
50 lb	W	-4593	7	2300	200	2.02	Adjusted
50 lb	X	-5328	-8	2300	200	2.02	Adjusted
50 lb	Y	-3738	2	2300	200	2.02	Adjusted

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist 10/01/2024 Dwight R Johnson, Reviewer 10/01/2024



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Certificate number: M25009
Calibration Date: 10/01/2024 Purchase Order Number:

Environmental conditions at time of test: Serial# F308

Temperature: 21.1 °C Humidity: 45.5 % Pressure: 675.6 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): **14 piece Metric Kit** SN F308

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	<i>k</i>	Condition As Left
2 kg		34	34	200	17	2.04	In-Tolerance
1 kg		29.0	29.0	100	8.7	2.04	In-Tolerance
500 g		7.5	7.5	70	6.1	2.04	In-Tolerance
200 g		10.6	10.6	40	3.4	2.04	In-Tolerance
200 g		13.4	13.4	40	3.4	2.04	In-Tolerance
100 g		6.8	6.8	20	1.7	2.04	In-Tolerance
50 g		2.08	2.08	10	0.86	2.04	In-Tolerance
20 g		1.37	1.37	4	0.34	2.04	In-Tolerance
20 g		0.61	0.61	4	0.34	2.04	In-Tolerance
10 g		0.80	0.80	2	0.17	2.04	In-Tolerance
5 g		0.16	0.16	1.5	0.13	2.04	In-Tolerance
2 g		0.936	0.936	1.1	0.095	2.04	In-Tolerance
2 g		0.716	0.716	1.1	0.095	2.04	In-Tolerance
1 g		0.272	0.272	0.9	0.078	2.04	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemed artifacts were tagged and must be placed out of service.
The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson *Dwight R Johnson*

Ron E Peterson, Metrologist 10/01/2024 Dwight R Johnson, Reviewer 10/01/2024



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Certificate number: M25009
 Calibration Date: 10/01/2024 Purchase Order Number:

Environmental conditions at time of test: Serial# 1190SD

Temperature: 21.1 °C Humidity: 45.5 % Pressure: 675.6 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): 20 piece Avoirdupois Kit SN 1190SD

Nominal	SN/ID	Correction as Found		Correction as Left		NIST Class F Tolerance (mg)	Uncertainty		Condition As Left
		mg		mg			mg	k	
5 lb	1	37		37		230	20	2.04	In-Tolerance
5 lb	2	101		101		230	20	2.04	In-Tolerance
5 lb	3	129		129		230	20	2.04	In-Tolerance
5 lb	4	16		16		230	20	2.04	In-Tolerance
5 lb	5	46		46		230	20	2.04	In-Tolerance
1 lb	1	30.5		30.5		70	6.1	2.04	In-Tolerance
1 lb	2	12.5		12.5		70	6.1	2.04	In-Tolerance
1 lb	3	16.5		16.5		70	6.1	2.04	In-Tolerance
1 lb	4	29.5		29.5		70	6.1	2.04	In-Tolerance
1 lb	5	35.5		35.5		70	6.1	2.04	In-Tolerance
8 oz		6.2		6.2		45	4.0	2.04	In-Tolerance
4 oz		3.8		3.8		23	2.0	2.03	In-Tolerance
2 oz		2.23		2.23		11	0.95	2.04	In-Tolerance
1 oz		2.27		2.27		5.4	0.48	2.03	In-Tolerance
0.5 oz		1.31		1.31		2.8	0.25	2.04	In-Tolerance
0.25 oz		1.19		1.19		1.7	0.15	2.03	In-Tolerance
0.125 oz		0.29		0.29		1.3	0.12	2.03	In-Tolerance
0.0625 oz		0.558		0.558		1.1	0.095	2.03	In-Tolerance
0.03125 oz		0.323		0.323		0.87	0.077	2.03	In-Tolerance
0.03125 oz	.	0.301		0.301		0.87	0.077	2.03	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist 10/01/2024 Dwight R Johnson, Reviewer 10/01/2024

Office of Weights and Measures
Metrology Laboratory

Office: 118 West Capitol Avenue, Pierre, SD 57501

Lab: 1100 Otter Rd, Bldg D, Sturgis, SD 57785

Lab: 605-347-7541, Office: 605-773-3697, Cell: 605-280-4572

Email: ron.peterson@state.sd.us <https://dps.sd.gov/inspections/weights-measures>

CALIBRATION CERTIFICATE

Capital Scale (Big Red)

SA# 61

Certificate number: M25048

Physical Address:

Billing Address:

3021 Valley Forge Street

3021 Valley Forge Street

Bismarck, ND 58503

Bismarck, ND 58503

Contact: Travis Will

Received Date: 01/23/2025

Phone: 701-255-1556

Certificate Issued: 01/27/2025

Artifacts Submitted and Summary of Results:

Quantity	Artifact	Total Pieces	Recvd in Tol	Adjusted	Rejected	As Left In Tolerance
2	2000 lb Weight Carts	2	0	2	0	2
16	1000 lb Weights	16	14	7	0	16
20	50 lb Weights	20	13	15	0	20
1	Avoirdupois Kit	20	20	0	0	20
1	Metric Kit	14	14	0	0	14

Uncertainty Statement: The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor *k* to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 2008 ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. For weight carts, factors included on the inspection checklist have not been included in the calibration uncertainty. However, factors on the checklist may contribute measurement errors that are significant if not properly maintained during use.

Conformity Statement:

The artifacts submitted for this calibration are calibrated to NIST Handbook 105-1 (1990 or 2019), NIST Handbook 105-8 (2019), NIST Handbook 105-3 (2010), or ASTM E617 (2023), Standard Specification for Laboratory Weights and Precision Mass Standards specifications. The reported test values relate only to the observations made at the time and conditions of the test. Artifacts fully comply with all requirements (both specifications and tolerances) of the applicable documentary standard unless otherwise stated. Stated expanded uncertainties are less than one-third of the specified tolerances (maximum permissible errors, m.p.e.) for mass calibrations and less than the specified tolerances for volume calibrations. The correction value plus or minus the expanded uncertainty is within the stated tolerances. It is the decision rule of the SD State Metrology Laboratory that any cast weights determined to have a correction within 66 % of the upper tolerance or 50 % of the lower tolerance will be adjusted closer to zero mass correction, even if the mass correction originally met the applicable tolerance.

Traceability Statement:

The Standards of the SD Metrology Laboratory used for comparison are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory certificate number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.



Ron E Peterson, Metrologist

01/27/2025



NVLAP LAB CODE 600384-0

Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 600384-0. This certificate may not be used to claim product endorsement by NVLAP, NIST Office of Weights and Measures or any other government agency, and may not be reproduced, except in full without written approval from this laboratory.



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: **Capital Scale (Big Red)**
 Calibration Date: **01/27/2025**

Certificate Number: **M25048**

Environmental conditions at time of test:

Temperature: 19.55 °C Humidity: 48.63 % Pressure: 666.88 mmhg

Test method used: SOP 33 Calibrations of Weight Carts, May 2019

Test equipment used: Recently calibrated weights and a Mettler SLS510 Load Cell with IND570 Indicator.
 Vaisala PT301

Condition of Carts: Used but in good condition

Manufacturer: Unk

SN: 541094

Nominal (lb)	AS Found (lb)	As Found (g)	As Left (lb)	As Left (g)	Uncertainty (lb)	k	Tolerance (lb)	Condition as Left
2000	2.40	1090	0.05	23	0.11	2.01	0.70	Adjusted

Notes:

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory.

The above weight cart was allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted if needed and as noted above to as close as practical to zero error. All fluid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require recalibration of the weight cart prior to subsequent use.

Conformity Assessment:

The weight cart identified on this calibration certificate complies with NIST Handbook 105-8, 2019 specifications and tolerances. Additional details regarding the assessment are included in the associated checklist that is an integral part of this calibration certificate. The weight cart was found (or adjusted) to within the specified tolerances.

The above weight cart was compared with standards of the State of South Dakota, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and have current calibration values. The assigned certificate number provides documented evidence for measurement traceability.

Ron E Peterson, Metrologist

01/27/2025

Ver 20250114



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



Inspection Checklist for Weight Cart

Calibrated for: Capital Scale (Big Red) Certificate number: M25048
 Calibration Date: 01/28/2025

Manufacturer: Date of Manufacture:
 Model Number: ID/SN Number:

Nominal Mass of Weight Cart: 2000 lbs Suitably marked: Yes/No
 Powered by: Electric/generator Diesel Gasoline
 Fluid Levels: Engine Oil
 Hydraulic Fluid Sealed: Yes/No
 Battery Sealed: Yes/No
 Liquid Fuel Reference Line Present: Yes/No

Fluid drain tubes extend beyond the body of the cart: Yes/No
 Number of axles:
 Number /Size of Tires:
 Sealed wheel bearings: Yes/No
 Drain holes present in locations where water may accumulate: Yes/No
 Weight restraint railing permanently fixed and solid: Yes/No
 Adjusting cavity accessible: Yes/No Approximate capacity:(lbs)
 Adjusting cavity sealed: Yes/No
 Service brakes functioning properly: Yes/No
 Parking brakes functioning properly: Yes/No
 Remote control functioning properly: Yes/No

General condition at time of calibration (note any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals).

List and report any repair and maintenance performed, parts replaced, etc., Leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc. Include any comments or changes since the last calibration.

Ron E Peterson

Ron E Peterson, Metrologist 01/27/2025
 Ver Ver 20250114



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: **Capital Scale (Big Red)** Certificate Number: **M25048**
 Calibration Date: **01/27/2025**

Environmental conditions at time of test:

Temperature: 19.45 °C Humidity: 48.83 % Pressure: 666.92 mmhg

Test method used: SOP 33 Calibrations of Weight Carts, May 2019

Test equipment used: Recently calibrated weights and a Mettler SLS510 Load Cell with IND570 Indicator.
 Vaisala PT301

Condition of Carts: Used but in good condition

Manufacturer: Unk

SN: Unk

Nominal (lb)	AS Found (lb)	As Found (g)	As Left (lb)	As Left (g)	Uncertainty (lb)	k	Tolerance (lb)	Condition as Left
2000	11.89	5398	0.05	23	0.11	2.01	0.70	Adjusted

Notes:

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory.

The above weight cart was allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted if needed and as noted above to as close as practical to zero error. All fluid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require recalibration of the weight cart prior to subsequent use.

Conformity Assessment:

The weight cart identified on this calibration certificate complies with NIST Handbook 105-8, 2019 specifications and tolerances. Additional details regarding the assessment are included in the associated checklist that is an integral part of this calibration certificate. The weight cart was found (or adjusted) to within the specified tolerances.

The above weight cart was compared with standards of the State of South Dakota, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and have current calibration values. The assigned certificate number provides documented evidence for measurement traceability.

Ron E Peterson, Metrologist

01/27/2025

Ver 20250114



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



Inspection Checklist for Weight Cart

Calibrated for: Capital Scale (Big Red) Certificate number: M25048
Calibration Date: 01/28/2025

Manufacturer: Date of Manufacture:
Model Number: ID/SN Number:

Nominal Mass of Weight Cart: 2000 lbs Suitably marked: Yes/No
 Powered by: Electric/generator Diesel Gasoline
 Fluid Levels: Engine Oil

Hydraulic Fluid Sealed: Yes/No
Battery Sealed: Yes/No
Liquid Fuel Reference Line Present: Yes/No

Fluid drain tubes extend beyond the body of the cart: Yes/No
 Number of axles:
 Number /Size of Tires:
 Sealed wheel bearings: Yes/No
 Drain holes present in locations where water may accumulate: Yes/No
 Weight restraint railing permanently fixed and solid: Yes/No
 Adjusting cavity accessible: Yes/No Approximate capacity:(lbs)
 Adjusting cavity sealed: Yes/No
 Service brakes functioning properly: Yes/No
 Parking brakes functioning properly: Yes/No
 Remote control functioning properly: Yes/No

General condition at time of calibration (note any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals).

List and report any repair and maintenance performed, parts replaced, etc., Leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc. Include any comments or changes since the last calibration.

Ron E Peterson

Ron E Peterson, Metrologist 01/27/2025
Ver Ver 20250114



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Big Red) **Certificate number:** M25048
Calibration Date: 01/27/2025 **Purchase Order Number:** 0

Environmental conditions at time of test:

Temperature: 19.48 °C **Humidity:** 47.5 % **Pressure:** 666.7 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019
Test equipment used: Lab standards traceable to the SI, an XPE604KMC balance, and a Vaisala PTU301
Condition of Weights: Cleaned and painted

Artifact(s): 16 - 1000 lb weights

Nominal	SN/ID	Correction as Found		Correction as Left		NIST Class F Tolerance (g)	Uncertainty		Condition As Left
		lb	g	lb	g		g	k	
1000 lb	13.10	-0.02	-9.8	-0.02	-9.8	45	4.7	2.0	In-Tolerance
1000 lb	13.11	-0.05	-21.4	-0.05	-21.4	45	4.7	2.0	In-Tolerance
1000 lb	13.12	-0.07	-32.7	0.00	0.0	45	4.7	2.0	Adjusted
1000 lb	13.13	-0.03	-15.8	-0.03	-15.8	45	4.7	2.0	In-Tolerance
1000 lb	13.18	-0.07	-29.9	0.00	0.0	45	4.7	2.0	Adjusted
1000 lb	13.19	-0.06	-27.9	0.00	0.1	45	4.7	2.0	Adjusted
1000 lb	13.20	-0.02	-7.0	-0.02	-7.0	45	4.7	2.0	In-Tolerance
1000 lb	13.20	-0.09	-39.9	0.00	0.0	45	4.7	2.0	Adjusted
1000 lb	13.21	-0.04	-16.2	-0.04	-16.2	45	4.7	2.0	In-Tolerance
1000 lb	13.22	-0.07	-30.7	0.00	-0.1	45	4.7	2.0	Adjusted
1000 lb	13.23	-0.02	-8.0	-0.02	-8.0	45	4.7	2.0	In-Tolerance
1000 lb	13.24	-0.07	-33.8	0.00	0.2	45	4.7	2.0	Adjusted
1000 lb	13.25	-0.04	-16.9	-0.04	-16.9	45	4.7	2.0	In-Tolerance
1000 lb	13.27	-0.05	-21.3	-0.05	-21.3	45	4.7	2.0	In-Tolerance
1000 lb	13.28	-0.03	-14.2	-0.03	-14.2	45	4.7	2.0	In-Tolerance
1000 lb	13.29	-0.09	-39.7	0.00	0.0	45	4.7	2.0	Adjusted

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson
 Ron E Peterson, Metrologist

01/27/2025



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Big Red)
Calibration Date: 01/27/2025

Certificate number: M25048
Purchase Order Number:

Environmental conditions at time of test:

Temperature: 20.2 °C Humidity: 46.3 % Pressure: 667.3 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): **20 50 lb weights**

Nominal	SN/ID	Correction as Found	Correction as Left	NIST Class F	Uncertainty		Condition As Left
		mg	mg	Tolerance (mg)	mg	k	
50 lb	03	-3318	57	2300	200	2.02	Adjusted
50 lb	09	-1823	7	2300	200	2.02	Adjusted
50 lb	19	-2138	77	2300	200	2.02	Adjusted
50 lb	29	-2533	-8	2300	200	2.02	Adjusted
50 lb	32	-1148	-1148	2300	200	2.02	In-Tolerance
50 lb	42	-1433	7	2300	200	2.02	Adjusted
50 lb	43	-1933	47	2300	200	2.02	Adjusted
50 lb	44	-2328	17	2300	200	2.02	Adjusted
50 lb	45	-1853	42	2300	200	2.02	Adjusted
50 lb	51	-1903	7	2300	200	2.02	Adjusted
50 lb	60	-2183	-23	2300	200	2.02	Adjusted
50 lb	11B	-2353	77	2300	200	2.02	Adjusted
50 lb	11C	-728	-728	2300	200	2.02	In-Tolerance
50 lb	A	-3333	7	2300	200	2.02	Adjusted
50 lb	B	-718	-718	2300	200	2.02	In-Tolerance
50 lb	K	-1848	47	2300	200	2.02	Adjusted
50 lb	L	-1368	47	2300	200	2.02	Adjusted
50 lb	N	-318	-318	2300	200	2.02	In-Tolerance
50 lb	P	-1233	-23	2300	200	2.02	Adjusted
50 lb	Y	-808	-808	2300	200	2.02	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist

01/27/2025



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Big Red) Certificate number: M25048
 Calibration Date: 01/28/2025 Purchase Order Number:

Environmental conditions at time of test:

Temperature: 21 °C Humidity: 45.5 % Pressure: 665.4 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): **20 piece Avoirdupois Kit** SN 010813A

Nominal	SN/ID	Correction as Found	Correction as Left	NIST Class F	Uncertainty	k	Condition As Left
		mg	mg	Tolerance (mg)	mg		
5 lb	1	67	67	230	20	2.04	In-Tolerance
5 lb	2	66	66	230	20	2.04	In-Tolerance
5 lb	3	65	65	230	20	2.04	In-Tolerance
5 lb	4	62	62	230	20	2.04	In-Tolerance
5 lb	5	64	64	230	20	2.04	In-Tolerance
1 lb	1	23.5	23.5	70	6.1	2.04	In-Tolerance
1 lb	2	22.5	22.5	70	6.1	2.04	In-Tolerance
1 lb	3	24.5	24.5	70	6.1	2.04	In-Tolerance
1 lb	4	16.5	16.5	70	6.1	2.04	In-Tolerance
1 lb	5	26.5	26.5	70	6.1	2.04	In-Tolerance
8 oz		21.2	21.2	45	4.0	2.04	In-Tolerance
4 oz		9.4	9.4	23	2.0	2.03	In-Tolerance
2 oz		3.52	3.52	11	0.95	2.04	In-Tolerance
1 oz		1.65	1.65	5.4	0.48	2.03	In-Tolerance
0.5 oz		0.98	0.98	2.8	0.25	2.04	In-Tolerance
0.25 oz		0.38	0.38	1.7	0.15	2.03	In-Tolerance
0.125 oz		0.61	0.61	1.3	0.12	2.03	In-Tolerance
0.0625 oz		0.173	0.173	1.1	0.095	2.03	In-Tolerance
0.03125 oz		0.218	0.218	0.87	0.077	2.03	In-Tolerance
0.03125 oz		0.323	0.323	0.87	0.077	2.03	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Big Red) Certificate number: M25048
Calibration Date: 01/28/2025 Purchase Order Number:

Environmental conditions at time of test:

Temperature: 21 °C Humidity: 45.5 % Pressure: 665.4 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): 14 piece Avoirdupois Kit SN 11905E

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	k	Condition As Left
2 kg	2	42	42	200	17	2.04	In-Tolerance
1 kg		33.0	33.0	100	8.7	2.04	In-Tolerance
500 g		27.5	27.5	70	6.1	2.04	In-Tolerance
200 g		13.3	13.3	40	3.4	2.04	In-Tolerance
200 g		17.9	17.9	40	3.4	2.04	In-Tolerance
100 g		9.9	9.9	20	1.7	2.04	In-Tolerance
50 g		5.18	5.18	10	0.86	2.04	In-Tolerance
20 g		2.70	2.70	4	0.34	2.04	In-Tolerance
20 g		0.99	0.99	4	0.34	2.04	In-Tolerance
10 g		0.42	0.42	2	0.17	2.04	In-Tolerance
5 g		0.65	0.65	1.5	0.13	2.04	In-Tolerance
2 g		0.711	0.711	1.1	0.095	2.04	In-Tolerance
2 g		0.151	0.151	1.1	0.095	2.04	In-Tolerance
1 g		0.342	0.342	0.9	0.078	2.04	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist 01/28/2025

Office of Weights and Measures

Metrology Laboratory

Office: 118 West Capitol Avenue, Pierre, SD 57501

Lab: 1100 Otter Rd, Bldg D, Sturgis, SD 57785

Lab: 605-347-7541, Office: 605-773-3697, Cell: 605-280-4572

Email: ron.peterson@state.sd.us <https://dps.sd.gov/inspections/weights-measures>

CALIBRATION CERTIFICATE

Capital Scale (Shop)

SA# 61

Certificate number: M25049

Physical Address:

Billing Address:

3021 Valley Forge Street

3021 Valley Forge Street

Bismarck, ND 58503

Bismarck, ND 58503

Contact: Travis Will

Received Date: 01/23/2025

Phone: 701-255-1556

Certificate Issued: 01/27/2025

Artifacts Submitted and Summary of Results:

Quantity	Artifact	Total Pieces	Recvd in Tol	Adjusted	Rejected	As Left In Tolerance
20	50 lb Weights	20	19	7	1	19
3	25 lb and 10 lb cast Weights	3	3	0	0	3
1	Metric Kit	21	21	0	0	21
2	Avoirdupois Kits	30	30	0	0	30

Uncertainty Statement: The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor *k* to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 2008 ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. For weight carts, factors included on the inspection checklist have not been included in the calibration uncertainty. However, factors on the checklist may contribute measurement errors that are significant if not properly maintained during use.

Conformity Statement:

The artifacts submitted for this calibration are calibrated to NIST Handbook 105-1 (1990 or 2019), NIST Handbook 105-8 (2019), NIST Handbook 105-3 (2010), or ASTM E617 (2023), Standard Specification for Laboratory Weights and Precision Mass Standards specifications. The reported test values relate only to the observations made at the time and conditions of the test. Artifacts fully comply with all requirements (both specifications and tolerances) of the applicable documentary standard unless otherwise stated. Stated expanded uncertainties are less than one-third of the specified tolerances (maximum permissible errors, m.p.e.) for mass calibrations and less than the specified tolerances for volume calibrations. The correction value plus or minus the expanded uncertainty is within the stated tolerances. It is the decision rule of the SD State Metrology Laboratory that any cast weights determined to have a correction within 66 % of the upper tolerance or 50 % of the lower tolerance will be adjusted closer to zero mass correction, even if the mass correction originally met the applicable tolerance.

Traceability Statement:

The Standards of the SD Metrology Laboratory used for comparison are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory certificate number identified above is the unique report number to be used in referencing measurement traceability for artifacts calibrated in this report only.



Ron E Peterson, Metrologist

01/27/2025



NVLAP LAB CODE 600384-0

Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 600384-0. This certificate may not be used to claim product endorsement by NVLAP, NIST Office of Weights and Measures or any other government agency, and may not be reproduced, except in full without written approval from this laboratory.



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Shop) Certificate number: M25049
 Calibration Date: 01/27/2025 Purchase Order Number:

Environmental conditions at time of test: Serial#

Temperature: 19.5 °C Humidity: 45.3 % Pressure: 666.4 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): **20 50 lb weights**

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	k	Condition As Left
50 lb	04	-1358	-18	2300	200	2.02	Adjusted
50 lb	11	-248	-248	2300	200	2.02	In-Tolerance
50 lb	13	72	72	2300	200	2.02	In-Tolerance
50 lb	15	-868	-868	2300	200	2.02	In-Tolerance
50 lb	18	-1208	17	2300	200	2.02	Adjusted
50 lb	20	-1023	-1023	2300	200	2.02	In-Tolerance
50 lb	21	762	762	2300	200	2.02	In-Tolerance
50 lb	24	-808	-808	2300	200	2.02	In-Tolerance
50 lb	33	-1008	-1008	2300	200	2.02	In-Tolerance
50 lb	34	-1213	2	2300	200	2.02	Adjusted
50 lb	34	-268	-268	2300	200	2.02	In-Tolerance
50 lb	37	32	32	2300	200	2.02	In-Tolerance
50 lb	39	-673	Broken Shoulder	2300	200	2.02	Reject
50 lb	40	-873	-873	2300	200	2.02	In-Tolerance
50 lb	70	-1958	27	2300	200	2.02	Adjusted
50 lb	71	-1193	2	2300	200	2.02	Adjusted
50 lb	72	-403	-403	2300	200	2.02	In-Tolerance
50 lb	77	-128	-128	2300	200	2.02	In-Tolerance
50 lb	11F	-1618	27	2300	200	2.02	Adjusted
50 lb	AAA	-1303	-13	2300	200	2.02	Adjusted

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist 01/27/2025



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Shop) Certificate number: M25049

Calibration Date: 01/24/2025 Purchase Order Number:

Environmental conditions at time of test: Serial#

Temperature: 21.2 °C Humidity: 45.5 % Pressure: 667.1 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): 3 Avoirdupois Weight(s)

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	k	Condition As Left
25 lb	.	-274	-274	1100	120	2.02	In-Tolerance
10 lb	S	235	235	450	39	2.04	In-Tolerance
10 lb	1	-28	-28	450	39	2.04	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service. The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Shop) Certificate number: M25049
 Calibration Date: 01/24/2025 Purchase Order Number:

Environmental conditions at time of test: Serial# 11111A

Temperature: 21 °C Humidity: 45 % Pressure: 667.3 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): **21 piece Metric Kit** SN 11111A

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	k	Condition As Left
5 kg	1	162	162	500	43	2.04	In-Tolerance
5 kg	2	142	142	500	43	2.04	In-Tolerance
2 kg	1	33	33	200	17	2.04	In-Tolerance
2 kg		44	44	200	17	2.04	In-Tolerance
1 kg		46.0	46.0	100	8.7	2.04	In-Tolerance
500 g	1	16.5	16.5	70	6.1	2.04	In-Tolerance
500 g	2	20.5	20.5	70	6.1	2.04	In-Tolerance
500 g	3	13.5	13.5	70	6.1	2.04	In-Tolerance
500 g	4	20.5	20.5	70	6.1	2.04	In-Tolerance
500 g	5	22.5	22.5	70	6.1	2.04	In-Tolerance
200 g	1	9.1	9.1	40	3.4	2.04	In-Tolerance
200 g	2	9.8	9.8	40	3.4	2.04	In-Tolerance
100 g		8.0	8.0	20	1.7	2.04	In-Tolerance
50 g		1.60	1.60	10	0.86	2.04	In-Tolerance
20 g		1.69	1.69	4	0.34	2.04	In-Tolerance
20 g		1.23	1.23	4	0.34	2.04	In-Tolerance
10 g		0.53	0.53	2	0.17	2.04	In-Tolerance
5 g		0.72	0.72	1.5	0.13	2.04	In-Tolerance
2 g		0.286	0.286	1.1	0.095	2.04	In-Tolerance
2 g		0.571	0.571	1.1	0.095	2.04	In-Tolerance
1 g		0.152	0.152	0.9	0.078	2.04	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight





CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Shop) Certificate number: M25049
Calibration Date: 01/24/2025 Purchase Order Number:

Environmental conditions at time of test: Serial# 11905B
Temperature: 21 °C Humidity: 45 % Pressure: 667.3 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019
Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301
Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): **17 piece Avoirdupois Kit** SN 11905B

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	k	Condition As Left
2 lb	1	27.4	27.4	91	7.9	2.04	In-Tolerance
2 lb	2	11.4	11.4	91	7.9	2.04	In-Tolerance
2 lb	3	12.4	12.4	91	7.9	2.04	In-Tolerance
1 lb		21.5	21.5	70	6.1	2.04	In-Tolerance
0.3 lb		-0.6	-0.6	27	3.9	2.04	In-Tolerance
0.2 lb		1.0	1.0	18	1.6	2.04	In-Tolerance
0.1 lb		4.32	4.32	9.1	0.78	2.04	In-Tolerance
0.05 lb		0.47	0.47	4.5	0.39	2.04	In-Tolerance
0.03 lb		-1.39	-1.39	2.7	0.39	2.04	In-Tolerance
0.02 lb		0.23	0.23	1.8	0.16	2.04	In-Tolerance
0.01 lb		-0.41	-0.41	1.5	0.13	2.04	In-Tolerance
8 oz		10.2	10.2	45	4.0	2.04	In-Tolerance
4 oz		-2.5	-2.5	23	2.0	2.03	In-Tolerance
2 oz		4.48	4.48	11	0.95	2.04	In-Tolerance
1 oz		0.62	0.62	5.4	0.48	2.03	In-Tolerance
0.5 oz		0.11	0.11	2.8	0.25	2.04	In-Tolerance
0.25 oz		0.37	0.37	1.7	0.15	2.03	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist 01/24/2025



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Shop) Certificate number: M25049
 Calibration Date: 01/24/2025 Purchase Order Number:
 Environmental conditions at time of test: Serial# 11905A

Temperature: 21 °C Humidity: 45 % Pressure: 667.3 mmHg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019
 Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301
 Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): **13 piece Avoirdupois Kit** SN 11905A

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	k	Condition As Left
10 lb		67	67	450	39	2.04	In-Tolerance
10 lb	.	99	99	450	39	2.04	In-Tolerance
5 lb		-92	-92	230	20	2.04	In-Tolerance
2 lb		34.4	34.4	91	7.9	2.04	In-Tolerance
2 lb	.	20.4	20.4	91	7.9	2.04	In-Tolerance
1 lb		-2.6	-2.6	70	6.1	2.04	In-Tolerance
8 oz		14.2	14.2	45	4.0	2.04	In-Tolerance
4 oz		-4.3	-4.3	23	2.0	2.03	In-Tolerance
1 oz		3.01	3.01	5.4	0.48	2.03	In-Tolerance
1 oz	.	3.01	3.01	5.4	0.48	2.03	In-Tolerance
1 oz	..	1.75	1.75	5.4	0.48	2.03	In-Tolerance
0.5 oz		1.53	1.53	2.8	0.25	2.04	In-Tolerance
0.25 oz		-1.29	-1.29	1.7	0.15	2.03	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist 01/24/2025

Office of Weights and Measures

Metrology Laboratory

Office: 118 West Capitol Avenue, Pierre, SD 57501

Lab: 1100 Otter Rd, Bldg D, Sturgis, SD 57785

Lab: 605-347-7541, Office: 605-773-3697, Cell: 605-280-4572

Email: ron.peterson@state.sd.us <https://dps.sd.gov/inspections/weights-measures>

CALIBRATION CERTIFICATE

Capital Scale (Trailer)

SA# 61

Certificate number: M25050

Physical Address:

Billing Address:

3021 Valley Forge Street

3021 Valley Forge Street

Bismarck, ND 58503

Bismarck, ND 58503

Contact: Travis Will

Received Date: 01/23/2025

Phone: 701-255-1556

Certificate Issued: 01/27/2025

Artifacts Submitted and Summary of Results:

Quantity	Artifact	Total Pieces	Recvd in Tol	Adjusted	Rejected	As Left In Tolerance
4	1000 lb Weights	3	4	2	0	4
30	50 lb Weights	30	26	12	0	30

Uncertainty Statement: The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor *k* to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 2008 ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. For weight carts, factors included on the inspection checklist have not been included in the calibration uncertainty. However, factors on the checklist may contribute measurement errors that are significant if not properly maintained during use.

Conformity Statement:

The artifacts submitted for this calibration are calibrated to NIST Handbook 105-1 (1990 or 2019), NIST Handbook 105-8 (2019), NIST Handbook 105-3 (2010), or ASTM E617 (2023), Standard Specification for Laboratory Weights and Precision Mass Standards specifications. The reported test values relate only to the observations made at the time and conditions of the test. Artifacts fully comply with all requirements (both specifications and tolerances) of the applicable documentary standard unless otherwise stated. Stated expanded uncertainties are less than one-third of the specified tolerances (maximum permissible errors, m.p.e.) for mass calibrations and less than the specified tolerances for volume calibrations. The correction value plus or minus the expanded uncertainty is within the stated tolerances. It is the decision rule of the SD State Metrology Laboratory that any cast weights determined to have a correction within 66 % of the upper tolerance or 50 % of the lower tolerance will be adjusted closer to zero mass correction, even if the mass correction originally met the applicable tolerance.

Traceability Statement:

The Standards of the SD Metrology Laboratory used for comparison are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory certificate number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.



Ron E Peterson, Metrologist

01/27/2025



NVLAP LAB CODE 600384-0

Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 600384-0. This certificate may not be used to claim product endorsement by NVLAP, NIST Office of Weights and Measures or any other government agency, and may not be reproduced, except in full without written approval from this laboratory.



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: **Capital Scale (Trailer)** Certificate number: **M25050**

Calibration Date: **01/24/2025** Purchase Order Number: **0**

Environmental conditions at time of test: Serial#

Temperature: **19.8 °C** Humidity: **44.6 %** Pressure: **667 mmhg**

Test method used: **SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019**

Test equipment used: **Lab standards traceable to the SI, an XPE604KMC balance, and a Vaisala PTU301**

Condition of Weights: **Cleaned and painted**

Artifact(s): 4 - 1000 lb weights

Nominal	SN/ID	Correction as Found		Correction as Left		NIST Class F Tolerance (g)	Uncertainty		Condition As Left
		lb	g	lb	g		g	k	
1000 lb	13.14	-0.03	-12.3	-0.03	-12.3	45	4.7	2.0	In-Tolerance
1000 lb	13.15	-0.06	-29.3	0.00	-0.1	45	4.7	2.0	Adjusted
1000 lb	13.16	-0.05	-22.9	0.00	0.0	45	4.7	2.0	Adjusted
1000 lb	13.17	-0.02	-11.1	-0.02	-11.1	45	4.7	2.0	In-Tolerance

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight



Ron E Peterson, Metrologist

01/24/2025



South Dakota Department of Public Safety
 Office of Weights and Measures
 Metrology Lab
 Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
 Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Trailer) Certificate number: M25050
 Calibration Date: 01/27/2025 Purchase Order Number:

Environmental conditions at time of test: Serial#
 Temperature: 20 °C Humidity: 48 % Pressure: 667.2 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019
 Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301
 Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): 20 50 lb weights

Nominal	SN/ID	Correction as Found	Correction as Left	NIST Class F	Uncertainty	k	Condition As Left
		mg	mg	Tolerance (mg)	mg		
50 lb	02	-923	-923	2300	200	2.02	In-Tolerance
50 lb	7	-1118	-1118	2300	200	2.02	In-Tolerance
50 lb	08	-498	-498	2300	200	2.02	In-Tolerance
50 lb	10	-1018	-1018	2300	200	2.02	In-Tolerance
50 lb	16	-1258	-8	2300	200	2.02	Adjusted
50 lb	27	-383	-383	2300	200	2.02	In-Tolerance
50 lb	30	-423	-423	2300	200	2.02	In-Tolerance
50 lb	30	-44448	2	2300	200	2.02	Adjusted
50 lb	31	-2248	17	2300	200	2.02	Adjusted
50 lb	35	-1163	-1163	2300	200	2.02	In-Tolerance
50 lb	36	-2213	37	2300	200	2.02	Adjusted
50 lb	41	-1078.000	-1078	2300	200	2.02	In-Tolerance
50 lb	46	-518.000	-518	2300	200	2.02	In-Tolerance
50 lb	48	-1048.000	-1048	2300	200	2.02	In-Tolerance
50 lb	49	-483.000	-483	2300	200	2.02	In-Tolerance
50 lb	50	-753.000	-753	2300	200	2.02	In-Tolerance
50 lb	54	-1523.000	12	2300	200	2.02	Adjusted
50 lb	55	-1068.000	-1068	2300	200	2.02	In-Tolerance
50 lb	61	-1148.000	-1148	2300	200	2.02	In-Tolerance
50 lb	64	-1318.000	32	2300	200	2.02	Adjusted

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist 01/27/2025



South Dakota Department of Public Safety
Office of Weights and Measures
Metrology Lab
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Trailer) Certificate number: M25050
 Calibration Date: 01/27/2025 Purchase Order Number:

Environmental conditions at time of test: Serial#

Temperature: 20 °C Humidity: 48 % Pressure: 667.2 mmHg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Suitable for use. No significant wear or damage

Artifact(s): **10 50 lb weights**

Nominal	SN/ID	Correction as Found		Correction as Left		NIST Class F Tolerance (mg)	Uncertainty		Condition As Left
		mg		mg			mg	k	
50 lb	.	-1948		22		2300	200	2.02	Adjusted
50 lb	A	907		907		2300	200	2.02	In-Tolerance
50 lb	B	-643		-643		2300	200	2.02	In-Tolerance
50 lb	E	-1943		27		2300	200	2.02	Adjusted
50 lb	J	-1473		-13		2300	200	2.02	Adjusted
50 lb	M	-2318		2		2300	200	2.02	Adjusted
50 lb	P	-1858		-13		2300	200	2.02	Adjusted
50 lb	R	-1048		-1048		2300	200	2.02	In-Tolerance
50 lb	T	-158		-158		2300	200	2.02	In-Tolerance
50 lb	U	-1418		17		2300	200	2.02	Adjusted

* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.
 The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist 01/27/2025