



## Office of Weights and Measures

### Metrology Laboratory

Office: 118 West Capitol Avenue, Pierre, SD 57501

Lab: 1100 Otter Rd, Bldg D, Sturgis, SD 57785

Lab: 605-347-7541, Office: 605-773-3697, Cell: 605-280-4572

Email: [ron.peterson@state.sd.us](mailto:ron.peterson@state.sd.us) <https://dps.sd.gov/inspections/weights-measures>

## CALIBRATION CERTIFICATE

Capital Scale (Wanda)

Physical Address:

3021 Valley Forge St  
Bismarck, ND 58503

Contact: Travis Will

Phone: 701-255-1556

Certificate number: M26023

Billing Address:

3021 Valley Forge St  
Bismarck, ND 58503

Received Date: 08/25/2025

Certificate Issued: 08/26/2025

### Artifacts Submitted and Summary of Results:

Quantity	Artifact	Total Pieces	Recv'd in Tol	Adjusted	Rejected	As Left In Tolerance
2	2000 lb Weight Carts	2	2	1	0	2
16	1000 lb Weights	16	16	0	0	16
20	50 lb weights	20	19	2	0	20
1	Avoirdupois kit	20	20	0	0	20
1	Metric kit	14	14	0	0	14

**Uncertainty Statement:** The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor  $k$  to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 2008 ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. For weight carts, factors included on the inspection checklist have not been included in the calibration uncertainty. However, factors on the checklist may contribute measurement errors that are significant if not properly maintained during use.

#### Conformity Statement:

The artifacts submitted for this calibration are calibrated to NIST Handbook 105-1 (1990 or 2019), NIST Handbook 105-8 (2019), NIST Handbook 105-3 (2010), or ASTM E617 (2023), Standard Specification for Laboratory Weights and Precision Mass Standards specifications. The reported test values relate only to the observations made at the time and conditions of the test. Artifacts fully comply with all requirements (both specifications and tolerances) of the applicable documentary standard unless otherwise stated. Stated expanded uncertainties are less than one-third of the specified tolerances (maximum permissible errors, m.p.e.) for mass calibrations and less than the specified tolerances for volume calibrations. The correction value plus or minus the expanded uncertainty is within the stated tolerances. It is the decision rule of the SD State Metrology Laboratory that any cast weights determined to have a correction within 66 % of the upper tolerance or 50 % of the lower tolerance will be adjusted closer to zero mass correction, even if the mass correction originally met the applicable tolerance.. SD Metrology laboratory uses an assumed density provided by the customer or weight manufacturer which could affect measurement results.

#### Traceability Statement:

The Standards of the SD Metrology Laboratory used for comparison are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory certificate number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

---

Ron E Peterson, Metrologist

08/26/2025

Wade Robbins, Reviewer

08/26/2025

---



Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 600384-0. This certificate may not be used to claim product endorsement by NVLAP, NIST Office of Weights and Measures or any other government agency, and may not be reproduced, except in full without written approval from this laboratory.



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Wanda)  
Calibration Date: 08/26/2025

Certificate Number: M26023

### Environmental conditions at time of test:

Temperature: 21.5 °C      Humidity: 51.2 %      Pressure: 672.4 mmhg

Test method used: SOP 33 Calibrations of Weight Carts, May 2019

Test equipment used: Recently calibrated weights and a Sartorius Loadcell PR6246/33 C6 with X4 Process controller .  
Vaisala PT301

Condition of Carts: Used but in good condition

Manufacturer: Unk

SN: 2016-1

Nominal (lb)	AS Found (lb)	As Found (g)	As Left (lb)	As Left (g)	Uncertainty (lb)	k	Tolerance (lb)	Condition as Left
2000	0.03	11	0.03	11	0.11	2.01	0.70	In-Tolerance

### Notes:

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory.

The above weight cart was allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted if needed and as noted above to as close as practical to zero error. All fluid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require recalibration of the weight cart prior to subsequent use.

### Conformity Assessment:

The weight cart identified on this calibration certificate complies with NIST Handbook 1058, 2019 specifications and tolerances. Additional details regarding the assessment are included in the associated checklist that is an integral part of this calibration certificate. The weight cart was found (or adjusted) to within the specified tolerances.

The above weight cart was compared with standards of the State of South Dakota, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and have current calibration values. The assigned certificate number provides documented evidence for measurement traceability.

Ron E Peterson, Metrologist

08/26/2025

Wade Robbins, Reviewer

08/26/2025

Ver 20250114



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



### Inspection Checklist for Weight Cart

Calibrated for: Capital Scale (Wanda)

Certificate number: M26023

Calibration Date: 08/27/2025

Manufacturer:

Unk
2000

Date of Manufacture

2016

Model Number:

ID/SN Number

2016-1

<input checked="" type="checkbox"/>	Nominal Mass of Weight Cart	2000 lbs	Suitably marked: Yes/No	Yes
<input checked="" type="checkbox"/>	Powered by:	Electric/generator	Diesel	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Fluid Levels:	Engine Oil	Gasoline	<input type="checkbox"/>
		Hydraulic Fluid		
		Battery		
		Liquid Fuel		
	Fluid drain tubes extend beyond the body of the cart:	Yes/No	Sealed: Yes/No	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Number of axles:	2	Sealed: Yes/No	Yes
<input checked="" type="checkbox"/>	Number /Size of Tires	15x5x11.25	Reference Line Present: Yes/No	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Sealed wheel bearings: Yes/No	Yes		
<input checked="" type="checkbox"/>	Drain holes present in locations where water may accumulate:	Yes/No		
<input checked="" type="checkbox"/>	Weight restraint railing permanently fixed and solid:	Yes/No		
<input checked="" type="checkbox"/>	Adjusting cavity accessible:	Yes	Yes	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Adjusting cavity sealed:	Yes	Yes	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Service brakes functioning properly:	Yes/No		
<input checked="" type="checkbox"/>	Parking brakes functioning properly:	Yes/No		
	Remote control functioning properly:	Yes/No		
	General condition at time of calibration (note any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals).	<input type="checkbox"/>		
<input checked="" type="checkbox"/>	List and report any repair and maintenance performed, parts replaced, etc., Leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc. Include any comments or changes since the last calibration.	<input type="checkbox"/>		

Ron E Peterson, Metrologist

08/26/2025

Wade Robbins, Reviewer

08/26/2025

Ver

Ver 20250114



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Wanda)  
Calibration Date: 08/26/2025

Certificate Number: M26023

### Environmental conditions at time of test:

Temperature: 21.46 °C      Humidity: 51.21 %      Pressure: 672.38 mmhg

Test method used: SOP 33 Calibrations of Weight Carts, May 2019

Test equipment used: Recently calibrated weights and a Sartorius Loadcell PR6246/33 C6 with X4 Process controller .  
Vaisala PT301

Condition of Carts: Used but in good condition

Manufacturer: UNK      SN: 2016-2

Nominal (lb)	AS Found (lb)	As Found (g)	As Left (lb)	As Left (g)	Uncertainty (lb)	k	Tolerance (lb)	Condition as Left
2000	0.18	80	-0.13	-61	0.11	2.01	0.70	Adjusted

### Notes:

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory.

The above weight cart was allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted if needed and as noted above to as close as practical to zero error. All fluid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require recalibration of the weight cart prior to subsequent use.

### Conformity Assessment:

The weight cart identified on this calibration certificate complies with NIST Handbook 105-8, 2019 specifications and tolerances. Additional details regarding the assessment are included in the associated checklist that is an integral part of this calibration certificate. The weight cart was found (or adjusted) to within the specified tolerances.

The above weight cart was compared with standards of the State of South Dakota, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and have current calibration values. The assigned certificate number provides documented evidence for measurement traceability.

*Ron E Peterson*      08/26/2025      *Wade Robbins*      08/26/2025  
Ron E Peterson, Metrologist      Wade Robbins, Reviewer  
Ver 20250114



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



### Inspection Checklist for Weight Cart

Calibrated for: Capital Scale (Wanda)

Certificate number: M26023

Calibration Date: 08/27/2025

Manufacturer:

UNK

Date of Manufacture

2016

Model Number:

2000

ID/SN Number

2016-2

<input checked="" type="checkbox"/>	Nominal Mass of Weight Cart	2000 lbs	Suitably marked: Yes/No	Yes
<input checked="" type="checkbox"/>	Powered by:	Electric/generator	Diesel	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Fluid Levels:	Engine Oil	Gasoline	<input type="checkbox"/>
		Hydraulic Fluid		
		Battery		
		Liquid Fuel		
	Fluid drain tubes extend beyond the body of the cart:	Yes/No	Sealed: Yes/No	<input type="checkbox"/>
	Number of axles:	2	Sealed: Yes/No	Yes
	Number /Size of Tires	15x5x11.25	Reference Line Present: Yes/No	<input type="checkbox"/>
	Sealed wheel bearings: Yes/No	Yes		
	Drain holes present in locations where water may accumulate:	Yes/No		
	Weight restraint railing permanently fixed and solid:	Yes/No		
	Adjusting cavity accessible: Yes/No	Yes	Yes	<input type="checkbox"/>
	Adjusting cavity sealed: Yes/No	Yes	Yes	<input type="checkbox"/>
	Service brakes functioning properly: Yes/No			
	Parking brakes functioning properly: Yes/No			
	Remote control functioning properly: Yes/No			
<input checked="" type="checkbox"/>	General condition at time of calibration (note any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals).			
<input checked="" type="checkbox"/>	List and report any repair and maintenance performed, parts replaced, etc., Leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc. Include any comments or changes since the last calibration.			

Ron E Peterson, Metrologist

08/26/2025

Wade Robbins, Reviewer

08/26/2025

Ver

Ver 20250114



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metroylogy Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: **Capital Scale (Wanda)**

**Certificate number:** M26023

Calibration Date: 08/26/2025

**Environmental conditions at time of test:**

Temperature: 22.3 °C

Humidity: 51 %

Pressure: 673.3 mmhg

**Test method used:** SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI, an XPE604 KMC balance, and a Vaisala PTU301

**Condition of Weights:** Cleaned and painted

### Artifact(s):

### 16 1000 lb weights

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

---

Ron E. Peterson, Metropolologist

08/26/2025

---

Wade Robbins Reviewer

08/26/2025



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab



## CALIBRATION CERTIFICATE

**Calibrated for:** Capital Scale (Wanda)  
**Calibration Date:** 08/26/2025

Certificate number: M26023

**Environmental conditions at time of test:**

Temperature: 22.93 °C      Humidity: 48.05 %      Pressure: 673.49 mmhg

Humidity: 48.05 %

Pressure: 673.49 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

**Condition of Weights:** Suitable for use. No significant wear or damage.

Artifact(s): 20 50 lb weights

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing:** Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

John E. D.

---

---

Ron E Peterson, Metrologist

08/26/2025

Wade Robbins, Reviewer

20/20/2020



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Wanda)

**Certificate number:** M26023

Calibration Date: 08/26/2025

Purchase Order Number:

**Environmental conditions at time of test:**

Temperature: 21.28 °C

Humidity: 45.22 %

Pressure: 763.49 mmhg

**Test method used:** SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019.

**Test equipment used:** Lab standards traceable to the SI. XPR6400313DC, XPR50035C, XPR226CDR, XPR36C, Vaisala PTI1301.

**Condition of Weights:** Suitable for use. No significant wear or damage.

Artifact(s): **20 piece Avordupois Kit** SN 119050

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing:** Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

John E. Hart

---

Ron E Peterson, Metrologist

08/26/2025

---

Wade Robbins, Reviewer

08/26/2025



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale (Wanda)

Certificate number: M26023

**Calibration Date:** 08/26/2025

Purchase Order Number:

**Environmental conditions at time of test:**

Temperature: 21.28 °C      Humidity: 45.22 %      Pressure: 673.49 mmHg

Used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution - May 2010

Test equipment used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Condition of Weights:** Suitable for use. No significant wear.

### Artifact(s):

#### 14 piece Metric Kit

SN E308

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing:** Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

*John E. D.*

---

Ron E Peterson, Metrologist

08/26/2025

---

Wade Robbins, Reviewer

08/26/2025



## Office of Weights and Measures

### Metrology Laboratory

Office: 118 West Capitol Avenue, Pierre, SD 57501

Lab: 1100 Otter Rd, Bldg D, Sturgis, SD 57785

Lab: 605-347-7541, Office: 605-773-3697, Cell: 605-280-4572

Email: ron.peterson@state.sd.us <https://dps.sd.gov/inspections/weights-measures>

## CALIBRATION CERTIFICATE

Capital Scale (Shop)

Certificate number: M26024

Physical Address:

3021 Valley Forge St

Bismarck, ND 58503

Billing Address:

3021 Valley Forge St

Bismarck, ND 58503

Contact: Travis Will

Phone: 701-255-1556

Received Date: 08/25/2025

Certificate Issued: 08/26/2025

### Artifacts Submitted and Summary of Results:

Quantity	Artifact	Total Pieces	Recv'd in Tol	Adjusted	Rejected	As Left In Tolerance
1	50 lb weight	1	0	1	0	1

**Uncertainty Statement:** The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor *k* to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 2008 ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. For weight carts, factors included on the inspection checklist have not been included in the calibration uncertainty. However, factors on the checklist may contribute measurement errors that are significant if not properly maintained during use.

**Conformity Statement:**

The artifacts submitted for this calibration are calibrated to NIST Handbook 105-1 (1990 or 2019), NIST Handbook 105-8 (2019), NIST Handbook 105-3 (2010), or ASTM E617 (2023), Standard Specification for Laboratory Weights and Precision Mass Standards specifications. The reported test values relate only to the observations made at the time and conditions of the test. Artifacts fully comply with all requirements (both specifications and tolerances) of the applicable documentary standard unless otherwise stated. Stated expanded uncertainties are less than one-third of the specified tolerances (maximum permissible errors, m.p.e.) for mass calibrations and less than the specified tolerances for volume calibrations. The correction value plus or minus the expanded uncertainty is within the stated tolerances. It is the decision rule of the SD State Metrology Laboratory that any cast weights determined to have a correction within 66 % of the upper tolerance or 50 % of the lower tolerance will be adjusted closer to zero mass correction, even if the mass correction originally met the applicable tolerance.. SD Metrology laboratory uses an assumed density provided by the customer or weight manufacturer which could affect measurement results.

**Traceability Statement:**

The Standards of the SD Metrology Laboratory used for comparison are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory certificate number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

Ron E Peterson, Metrologist

08/26/2025

Wade Robbins, Reviewer

08/26/2025



Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 600384-0. This certificate may not be used to claim product endorsement by NVLAP, NIST Office of Weights and Measures or any other government agency, and may not be reproduced, except in full without written approval from this laboratory.



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

**Calibrated for:** Capital Scale (Shop)  
**Calibration Date:** 08/26/2025

Certificate number: M26024  
Purchase Order Number:

**Environmental conditions at time of test:**

Temperature: 22.93 °C      Humidity: 48.05 %      Pressure: 673.49 mmhg

Humidity: 48.05 %

Pressure: 673.49 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI XPR640031 D5C XPR50035C XPR2262BC XPR36C Viciplus RTU301

**Condition of Weights:** Suitable for use. No significant wear or damage.

Artifact(s): 1. 50 lb weights

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing:** Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

---

---

---

Ron F Peterson, Metrologist

08/26/2025

---

Wade Robbins, Reviewer

08/26/2025



### Office of Weights and Measures

#### Metrology Laboratory

Office: 118 West Capitol Avenue, Pierre, SD 57501

Lab: 1100 Otter Rd, Bldg D, Sturgis, SD 57785

Lab: 605-347-7541, Office: 605-773-3697, Cell: 605-280-4572

Email: ron.peterson@state.sd.us <https://dps.sd.gov/inspections/weights-measures>

## CALIBRATION CERTIFICATE

Capital Scale Systems

SA# 61

Certificate number: M26066

Physical Address:

Billing Address:

3021 Valley Forge Dr.

3021 Valley Forge Dr.

Bismarck, ND 58503

Bismarck, ND 58503

Contact: Travis Will

Received Date: 01/14/2026

Phone: 701-255-1556

Certificate Issued: 01/15/2026

#### Artifacts Submitted and Summary of Results:

Quantity	Artifact	Total Pieces	Recv'd in Tol	Adjusted	Rejected	As Left In Tolerance
2	4000 lb Weight Carts	2	0	2	0	2
20	1000 lb Weights	20	20	3	0	20
76	50 lb Weights	75	53	45	1	75
6	Cast Weights	6	5	3	0	6
24	1 lb Weights	24	20	0	4	20
3	Avoirdupois kits	50	50	0	0	50
2	Metric Kits	31	31	0	0	31

**Uncertainty Statement:** The combined standard uncertainty includes the standard uncertainty reported for the standard and the standard uncertainty for the measurement process. The combined standard uncertainty is multiplied by a coverage factor *k* to provide an expanded uncertainty which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 2008 ISO/IEC Guide to the Expression of Uncertainty in Measurement. The expanded uncertainty is not to be confused with a tolerance limit for the user during application. For weight carts, factors included on the inspection checklist have not been included in the calibration uncertainty. However, factors on the checklist may contribute measurement errors that are significant if not properly maintained during use.

#### Conformity Statement:

The artifacts submitted for this calibration are calibrated to NIST Handbook 105-1 (1990 or 2019), NIST Handbook 105-8 (2019), NIST Handbook 105-3 (2010), or ASTM E617 (2023), Standard Specification for Laboratory Weights and Precision Mass Standards specifications. The reported test values relate only to the observations made at the time and conditions of the test. Artifacts fully comply with all requirements (both specifications and tolerances) of the applicable documentary standard unless otherwise stated. Stated expanded uncertainties are less than one-third of the specified tolerances (maximum permissible errors, m.p.e.) for mass calibrations and less than the specified tolerances for volume calibrations. The correction value plus or minus the expanded uncertainty is within the stated tolerances. It is the decision rule of the SD State Metrology Laboratory that any cast weights determined to have a correction within 66 % of the upper tolerance or 50 % of the lower tolerance will be adjusted closer to zero mass correction, even if the mass correction originally met the applicable tolerance.. SD Metrology laboratory uses an assumed density provided by the customer or weight manufacturer which could affect measurement results.

#### Traceability Statement:

The Standards of the SD Metrology Laboratory used for comparison are traceable to the International System of Units (SI) through the National Institute of Standards and Technology. The laboratory certificate number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

Ron E Peterson, Metrologist

01/15/2026

Wade Robbins, Reviewer

01/15/2026



NVLAP LAB CODE 600384-0

Accredited by the National Voluntary Laboratory Accreditation Program for the specific scope of accreditation under lab code 600384-0. This certificate may not be used to claim product endorsement by NVLAP, NIST Office of Weights and Measures or any other government agency, and may not be reproduced, except in full without written approval from this laboratory.



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

Certificate Number: M26066

Calibration Date: 01/15/2026

### Environmental conditions at time of test:

Temperature: 21.12 °C

Humidity: 44.57 %

Pressure: 664.79 mmhg

Test method used: SOP 33 Calibrations of Weight Carts, May 2019

Test equipment used: Recently calibrated weights and a Mettler SLS510 Load Cell with IND570 Indicator.

Vaisala PT301

Condition of Carts: Used but in good condition

Manufacturer: Unk

SN: Unk

Nominal (lb)	AS Found (lb)	As Found (g)	As Left (lb)	As Left (g)	Uncertainty (lb)	k	Tolerance (lb)	Condition as Left
2000	0.90	408	-0.02	-9	0.11	2.01	0.70	Adjusted

### Notes:

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory.

The above weight cart was allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted if needed and as noted above to as close as practical to zero error. All fluid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require recalibration of the weight cart prior to subsequent use.

### Conformity Assessment:

The weight cart identified on this calibration certificate complies with NIST Handbook 1058, 2019 specifications and tolerances. Additional details regarding the assessment are included in the associated checklist that is an integral part of this calibration certificate. The weight cart was found (or adjusted) to within the specified tolerances.

The above weight cart was compared with standards of the State of South Dakota, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and have current calibration values. The assigned certificate number provides documented evidence for measurement traceability.

Ron E Peterson, Metrologist  
Ver 20260109

01/15/2026

Wade Robbins, Reviewer

01/15/2026



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



### Inspection Checklist for Weight Cart

Calibrated for: Capital Scale Systems Certificate number: M26066  
Calibration Date: 01/15/2026

Manufacturer:	Unk	Date of Manufacture	Unk
Model Number:	Unk	ID/SN Number	Unk

<input checked="" type="checkbox"/>	Nominal Mass of Weight Cart	2000 lbs	Suitably marked: Yes/No	Yes
<input checked="" type="checkbox"/>	Powered by:	Electric/generator	Diesel	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Fluid Levels:	Engine Oil	Gasoline	<input type="checkbox"/>

Hydraulic Fluid  
Battery  
Liquid Fuel

<input type="checkbox"/>	Sealed: Yes/No	Yes
<input type="checkbox"/>	Sealed: Yes/No	Yes

<input type="checkbox"/>	Reference Line Present: Yes/No	Yes
--------------------------	--------------------------------	-----

<input checked="" type="checkbox"/>	Fluid drain tubes extend beyond the body of the cart: Yes/No	Yes
<input checked="" type="checkbox"/>	Number of axles:	2
<input checked="" type="checkbox"/>	Number /Size of Tires	16.25x5x11
<input checked="" type="checkbox"/>	Sealed wheel bearings: Yes/No	Yes
<input checked="" type="checkbox"/>	Drain holes present in locations where water may accumulate: Yes/No	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Weight restraint railing permanently fixed and solid: Yes/No	Yes
<input checked="" type="checkbox"/>	Adjusting cavity accessible: Yes/No	Yes
<input checked="" type="checkbox"/>	Adjusting cavity sealed: Yes/No	Yes
<input checked="" type="checkbox"/>	Service brakes functioning properly: Yes/No	Yes
<input checked="" type="checkbox"/>	Parking brakes functioning properly: Yes/No	Yes
<input checked="" type="checkbox"/>	Remote control functioning properly: Yes/No	

Approximate capacity:(lbs) 25

<input checked="" type="checkbox"/>	General condition at time of calibration (note any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals).
-------------------------------------	--

<input checked="" type="checkbox"/>	List and report any repair and maintenance performed, parts replaced, etc., Leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc. Include any comments or changes since the last calibration.
-------------------------------------	--

Ron E Peterson, Metrologist

01/15/2026

Wade Robbins, Reviewer

01/15/2026

Ver

Ver 20260109



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

Certificate Number: M26066

Calibration Date: 01/15/2026

### Environmental conditions at time of test:

Temperature: 21.12 °C

Humidity: 44.57 %

Pressure: 664.79 mmhg

Test method used: SOP 33 Calibrations of Weight Carts, May 2019

Test equipment used: Recently calibrated weights and a Mettler SLS510 Load Cell with IND570 Indicator.

Vaisala PT301

Condition of Carts: Used but in good condition

Manufacturer: Unk

SN:

541094

Nominal (lb)	AS Found (lb)	As Found (g)	As Left (lb)	As Left (g)	Uncertainty (lb)	k	Tolerance (lb)	Condition as Left
2000	-0.89	-406	-0.04	-17	0.11	2.01	0.70	Adjusted

### Notes:

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory.

The above weight cart was allowed to come to environmental equilibrium in the laboratory prior to calibration. The weight cart was adjusted if needed and as noted above to as close as practical to zero error. All fluid levels must be maintained as close to reference levels as possible during use. Any maintenance, repairs or damage to weight cart or its components will likely result in an out-of-tolerance condition; therefore, maintenance or replacement of components such as batteries, tires, filters, etc. will require recalibration of the weight cart prior to subsequent use.

### Conformity Assessment:

The weight cart identified on this calibration certificate complies with NIST Handbook 105-8, 2019 specifications and tolerances. Additional details regarding the assessment are included in the associated checklist that is an integral part of this calibration certificate. The weight cart was found (or adjusted) to within the specified tolerances.

The above weight cart was compared with standards of the State of South Dakota, which are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and have current calibration values. The assigned certificate number provides documented evidence for measurement traceability.

Ron E Peterson, Metrologist  
Ver 20260109

01/15/2026

Wade Robbins, Reviewer

01/15/2026



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



### Inspection Checklist for Weight Cart

Calibrated for: Capital Scale Systems Certificate number: M26066  
Calibration Date: 01/15/2026

Manufacturer:	Unk	Date of Manufacture	Unk
Model Number:	Unk	ID/SN Number	541094

<input checked="" type="checkbox"/>	Nominal Mass of Weight Cart	2000 lbs	Suitably marked: Yes/No	Yes
<input checked="" type="checkbox"/>	Powered by:	Electric/generator	Diesel	<input type="checkbox"/>
<input checked="" type="checkbox"/>	Fluid Levels:	Engine Oil	Gasoline	<input type="checkbox"/>
		Hydraulic Fluid		
		Battery		
		Liquid Fuel		
	Fluid drain tubes extend beyond the body of the cart:	Yes/No	Sealed: Yes/No	<input type="checkbox"/>
	Number of axles:	2	Sealed: Yes/No	<input type="checkbox"/>
	Number /Size of Tires	16.25x5x11	Reference Line Present: Yes/No	<input type="checkbox"/>
	Sealed wheel bearings: Yes/No	Yes		
	Drain holes present in locations where water may accumulate: Yes/No		Yes	<input type="checkbox"/>
	Weight restraint railing permanently fixed and solid: Yes/No		Yes	<input type="checkbox"/>
	Adjusting cavity accessible: Yes/No	Yes	Approximate capacity:(lbs)	25
	Adjusting cavity sealed: Yes/No	Yes		
	Service brakes functioning properly: Yes/No	Yes		
	Parking brakes functioning properly: Yes/No	Yes		
	Remote control functioning properly: Yes/No			

<input checked="" type="checkbox"/>	General condition at time of calibration (note any accumulated dirt/debris, damage, loose parts, or evidence of tampering or unauthorized entry of seals).
<input checked="" type="checkbox"/>	List and report any repair and maintenance performed, parts replaced, etc., Leaks repaired, new battery, carburetor, exhaust system, wheels changed, welding performed, etc. Include any comments or changes since the last calibration.

Ron E Peterson, Metrologist

01/15/2026

Wade Robbins, Reviewer

01/15/2026

Ver

Ver 20260109



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: **Capital Scale Systems**

**Certificate number:** M26066

**Calibration Date:** 018/15/2025

Purchase Order Number: 0

**Environmental conditions at time of test:**

Serial#

Temperature: 665.1 °C      Humidity: 21 %      Pressure: 45 mmhg

Humidity: 21 %

Pressure: 45 mmhg

**Test method used:** SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI, an XPE604KMC balance, and a Vaisala PTU301

**Condition of Weights:** Cleaned and painted

Artifact(s): 20 1000 lb weights

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight.**

*Walter H. D.*

Wanda Robbins

---

---

Ron E Peterson, Metrologist

018/15/2025

---

Wade Robbins, Reviewer

018/15/2025



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

Certificate number: M26066

Calibration Date: 01/15/2026

Purchase Order Number:

Environmental conditions at time of test:

Temperature: 20.7 °C      Humidity: 45.8 %      Pressure: 666 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Cleaned and painted

Artifact(s): 28 50 lb weights

Nominal	SN/ID	Correction as Found mg	Correction as Left mg	NIST Class F Tolerance (mg)	Uncertainty mg	k	Condition As Left
50 lb	2	-2318	-18	2300	200	2.02	Adjusted
50 lb	3	-1363	-3	2300	200	2.02	Adjusted
50 lb	4	-28	-28	2300	200	2.02	In-Tolerance
50 lb	6	-118	-118	2300	200	2.02	In-Tolerance
50 lb	7	-743	-743	2300	200	2.02	In-Tolerance
50 lb	8	-1698	27	2300	200	2.02	Adjusted
50 lb	9	-2068	2	2300	200	2.02	Adjusted
50 lb	10	-2658	-3	2300	200	2.02	Adjusted
50 lb	13	-8	-8	2300	200	2.02	In-Tolerance
50 lb	15	-828	-828	2300	200	2.02	In-Tolerance
50 lb	16	-3003	27	2300	200	2.02	Adjusted
50 lb	18	-963	-963	2300	200	2.02	In-Tolerance
50 lb	19	-1193	-18	2300	200	2.02	Adjusted
50 lb	21	1257	1257	2300	200	2.02	In-Tolerance
50 lb	22	1137	1137	2300	200	2.02	In-Tolerance
50 lb	23	-843	-843	2300	200	2.02	In-Tolerance
50 lb	26	1667	27	2300	200	2.02	Adjusted
50 lb	27	-1593	27	2300	200	2.02	Adjusted
50 lb	29	-2123	22	2300	200	2.02	Adjusted
50 lb	30	-1518	12	2300	200	2.02	Adjusted
50 lb	31	-943	-943	2300	200	2.02	In-Tolerance
50 lb	32	-2748	-8	2300	200	2.02	Adjusted
50 lb	33	-1038	-1038	2300	200	2.02	In-Tolerance
50 lb	34	-783	-783	2300	200	2.02	In-Tolerance
50 lb	34	-73	-73	2300	200	2.02	In-Tolerance
50 lb	35	-1808	-28	2300	200	2.02	Adjusted
50 lb	36	-1328	32	2300	200	2.02	Adjusted
50 lb	37	182	182	2300	200	2.02	In-Tolerance

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist

01/15/2026

Wade Robbins, Reviewer

01/15/2026



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

Certificate number: M26066

Calibration Date: 01/15/2026

Purchase Order Number:

Environmental conditions at time of test:

Temperature: 20.7 °C

Humidity: 45.8 %

Pressure: 666 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

Test equipment used: Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

Condition of Weights: Cleaned and painted

Artifact(s):

26 50 lb weights

Nominal	SN/ID	Correction as Found	Correction as Left	NIST Class F	Uncertainty	k	Condition As Left
		mg	mg				
50 lb	40	-2238	-3	2300	200	2.02	Adjusted
50 lb	41	-1863	-3	2300	200	2.02	Adjusted
50 lb	42	-1763	12	2300	200	2.02	Adjusted
50 lb	43	-1968	-3	2300	200	2.02	Adjusted
50 lb	44	-2873	-18	2300	200	2.02	Adjusted
50 lb	45	-2508	7	2300	200	2.02	Adjusted
50 lb	46	-1718	-3	2300	200	2.02	Adjusted
50 lb	47	-2013	2	2300	200	2.02	Adjusted
50 lb	48	-2113	-18	2300	200	2.02	Adjusted
50 lb	49	-568	-568	2300	200	2.02	In-Tolerance
50 lb	50	-2973	17	2300	200	2.02	Adjusted
50 lb	50	-368	-368	2300	200	2.02	In-Tolerance
50 lb	51	-67538	2	2300	200	2.02	Adjusted
50 lb	52	-1363	47	2300	200	2.02	Adjusted
50 lb	53	-233	-233	2300	200	2.02	In-Tolerance
50 lb	54	-513	-513	2300	200	2.02	In-Tolerance
50 lb	55	-1933	-13	2300	200	2.02	Adjusted
50 lb	60	-1478	12	2300	200	2.02	Adjusted
50 lb	61	-1793	27	2300	200	2.02	Adjusted
50 lb	64	-593	-593	2300	200	2.02	In-Tolerance
50 lb	69	-473	-473	2300	200	2.02	In-Tolerance
50 lb	71	292	292	2300	200	2.02	In-Tolerance
50 lb	72	-213	-213	2300	200	2.02	In-Tolerance
50 lb	72	-103	-103	2300	200	2.02	In-Tolerance
50 lb	73	-718	-718	2300	200	2.02	In-Tolerance
50 lb	77	-53	-53	2300	200	2.02	In-Tolerance

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight

Ron E Peterson, Metrologist

01/15/2026

Wade Robbins, Reviewer

01/15/2026



South Dakota Department of Public Safety

## Office of Weights and Measures

## Weights and Measures

Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

Certificate number: M26066

**Calibration Date:** 01/15/2026

**Purchase Order Number:**

**Environmental conditions at time of test:**

Temperature: 20.7 °C      Humidity: 45.8 %      Pressure: 666 mmhg

Humidity: 45.8 %

Pressure: 666 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI. XPR640031DSC, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301.

**Condition of Weights:** Cleaned and painted

Artifact(s): 22 50 lb weights

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight.**

Walter H. L.

Wade Robbie

---

---

Ron E Peterson, Metrologist

01/15/2026

---

Wade Robbins, Reviewer

01/15/2026



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

Certificate number: M26066

**Calibration Date:** 01/15/2026

**Purchase Order Number:**

**Environmental conditions at time of test:**

Temperature: 20.7 °C

Humidity: 45.8 %

Pressure: 666 mmhg

**Test method used:** SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

**Condition of Weights: Cleaned and painted**

Artifact(s): 6 Avoirdupois Weight(s)

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing:** Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

Howard P. Johnson

Wade Robbie

---

---

### Ron E Peterson, Metrologist

01/15/2026

Wade Robbins, Reviewer

01/15/2026



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

**Certificate number:** M26066

**Calibration Date:** 01/15/2026

**Purchase Order Number:**

**Environmental conditions at time of test:**

Temperature: 21.4 °C

Humidity: 43.8 %

Pressure: 666.3 mmhg

**Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019**

**Test equipment used:** Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

**Condition of Weights:** Cleaned and painted

### Artifact(s):

## 24 Avoirdupois Weight(s)

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight.**

*James M. H. L.*

the lab overnight

---

---

Ron E Peterson, Metrologist

01/15/2026

---

Wade Robbins, Reviewer

01/15/2026





South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab



## CALIBRATION CERTIFICATE

**Calibrated for:** Capital Scale Systems  
**Calibration Date:** 01/15/2026

**Certificate number:** M26066

**Environmental conditions at time of test:**

Temperature: 21.6 °C      Humidity: 44.4 %      Pressure: 666.2 mmhg

Humidity: 44.4 %

Pressure: 666.2 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI. XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

**Condition of Weights:** Suitable for use. No significant wear or damage.

Artifact(s): **21 piece Metric Kit** SN 11111A

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing:** Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

*John E. Miller*

Wade Robbins

---

---

Ron F Peterson, Metrologist

01/15/2026

---

Wade Robbins, Reviewer

01/15/2026



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

**Certificate number:** M26066

**Calibration Date:** 01/15/2026

**Purchase Order Number:**

**Environmental conditions at time of test:**

Temperature: 21.5 °C

Pressure: 666.6 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226C, XPR36C, Vaisala PTU301

**Condition of Weights:** Suitable for use. No significant wear or damage.

### Artifact(s):

### 14 piece Metric Kit

SN 1190EE

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

Shane P. L.

Wade Robbins

---

---

## Ron E Peterson, Metrologist

01/15/2026

Wade Robbins, Reviewer

01/15/2026



**South Dakota Department of Public Safety**  
**Office of Weights and Measures**  
**Metrology Lab**



## CALIBRATION CERTIFICATE

**Calibrated for:** Capital Scale Systems  
**Calibration Date:** 01/15/2026

Certificate number: M26066

**Environmental conditions at time of test:**

Temperature: 21.65 °C      Humidity: 46.2 %      Pressure: 666.7 mmhg

Humidity: 46.2 %

Pressure: 666.7 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI. XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301.

**Condition of Weights:** Suitable for use. No significant wear or damage.

Artifact(s): **17 piece Avordunois Kit** SN: **11905B**

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing:** Thermal equilibrium was obtained by placing the artifacts in the lab overnight.

*James M. H. L.*

the lab overnight

---

---

Ron E Peterson, Metrologist

01/15/2026

---

Wade Robbins, Reviewer

01/15/2026



South Dakota Department of Public Safety  
Office of Weights and Measures  
Metrology Lab  
Lab: 1100 Otter Rd, Bldg. D Sturgis, SD 57785 Phone: 605-347-7541  
Office: 118 West Capitol Avenue Pierre, SD 57501 Phone: 605-773-3697



## CALIBRATION CERTIFICATE

Calibrated for: Capital Scale Systems

Certificate number: M26066

**Calibration Date:** 01/15/2026

**Purchase Order Number:**

**Environmental conditions at time of test:**

Temperature: 21.5 °C

Humidity: 46.05 %

**Pressure:** 666.5 mmhg

Test method used: SOP 8 Medium Accuracy Calibrations of Mass Standards by Modified Substitution, May 2019

**Test equipment used:** Lab standards traceable to the SI, XPR64003LD5C, XPR5003SC, XPR226CDR, XPR36C, Vaisala PTU301

**Condition of Weights:** Suitable for use. No significant wear or damage.

Artifact(s): **13 piece Avoirdupois Kit** SN 11905A

\* Adjusted artifacts are in tolerance. Rejected and Condemned artifacts were tagged and must be placed out of service.

The values reported relate only to those observations made at the time and conditions of the test. This calibration certificate, so numbered, may not be reproduced, except in full, without approval of the laboratory. These weights were not screened for magnetism or checked for density, and effects of magnetism or density are not included in the uncertainties.

**Treatment of artifacts prior to testing: Thermal equilibrium was obtained by placing the artifacts in the lab overnight.**



Wade Robbins

---

---

### Ron E Peterson, Metrologist

01/15/2026

Wade Robbins, Reviewer

01/15/2026